



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/840,200

05/05/2004

Karl Sieradzki

AZMED.0400

7238

39602 7590 02/21/2008  
NOBLITT & GILMORE, LLC.  
4800 NORTH SCOTTSDALE ROAD  
SUITE 6000  
SCOTTSDALE, AZ 85251

EXAMINER

HOUSTON, ELIZABETH

ART UNIT

PAPER NUMBER

3731

MAIL DATE

DELIVERY MODE

02/21/2008

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b> 10/840,200	<b>Applicant(s)</b> SIERADZKI ET AL.	
	<b>Examiner</b> ELIZABETH HOUSTON	<b>Art Unit</b> 3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 12 November 2007.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-28 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-28 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)            | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)   | Paper No(s)/Mail Date. _____                                      |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>04/14/05</u> .  | 6) <input type="checkbox"/> Other: _____                          |

## **DETAILED ACTION**

### ***Election/Restrictions***

1. Applicant's election without traverse of Group I, apparatus, claims 1-28, in the reply filed on 11/12/07 is acknowledged.

### ***Claim Objections***

2. Claims 11 and 12 are objected to because of the following informalities: The preamble is inconsistent with the preamble of the claim from which it depends. To maintain consistency throughout the claims, it is suggested that the preamble be change to "A medical device according to claim 1 comprising ...". Appropriate correction is required.

### ***Claim Rejections - 35 USC § 112***

3. The following is a quotation of the second paragraph of 35 U.S.C. 112:  

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.
4. Claim 12 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. Claim 12 recites the limitation "the surface" in line 1. There is insufficient antecedent basis for this limitation in the claim. Therefore the limitation "wherein the surface comprises a surface of the porous portion" is unclear in terms of what surface is comprising the surface of the porous portion.

***Claim Rejections - 35 USC § 102***

5. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

**6. Claims 1-4, 6-9, 11-14, 16-19, 21, 22, 24-27 are rejected under 35**

**U.S.C. 102(e) as being anticipated by Lye et al (US 7,294,409).**

7. Lye discloses a medical device (10) having a main structure (14) that is a stent and a rigid porous portion (12) that is a metal. The porous portion is attached or integrated to the main structure (Col 4, L 1-17). There is a drug maintained in the porous portion (Col 2, L 19-20). The pores have different sizes (Col 4, L 58-61). The main structure is stainless steel (C 4, L 43) and the porous portion includes gold and silver (C 2, L 64 - C 3, L5).

***Claim Rejections - 35 USC § 103***

8. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Art Unit: 3731

9. The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

10. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

**11. Claim 5, 10, 15, 20, 23 and 28 are rejected under 35 U.S.C. 103(a) as being unpatentable over Lye et al (US 7,294,409) in view of Armini (US 5,919,126).**

12. Lye discloses all the limitations substantially as claimed as stated above except for the adhesive layer including chromium.

13. Armini discloses a stent with a layer of gold coating and an adhesive layer of chromium in order to facilitate adhesion of the gold to a stainless steel stent (Abstract; C 1, L 19-21).

14. It would have been obvious to one having ordinary skill in the art at the time of the invention to choose the material of Chromium as a transition layer between a stainless steel base structure and a gold coating since it is old and well known in the art Chromium facilitates adhesion of gold to stainless steel as taught by Armini.

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to ELIZABETH HOUSTON whose telephone number is (571)272-7134. The examiner can normally be reached on M-F 9:00-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Todd Manahan can be reached on 571-272-4713. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/E. H./  
Examiner, Art Unit 3731

Application/Control Number: 10/840,200

Page 6

Art Unit: 3731

/Todd E Manahan/

Supervisory Patent Examiner, Art Unit 3731